

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065603.D
 Acq On : 26 Jan 2021 14:49
 Operator : JC/MD
 Sample : M1221-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 27 03:28:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	208627	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	360516	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	312937	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	109050	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	134600	49.04	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.08%
35) Dibromofluoromethane	7.550	113	105991	49.61	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.22%
50) Toluene-d8	10.058	98	450412	53.63	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.26%
62) 4-Bromofluorobenzene	12.373	95	148362	49.06	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
Data File : VN065603.D
Acq On : 26 Jan 2021 14:49
Operator : JC/MD
Sample : M1221-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 27 03:28:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

