

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065608.D
 Acq On : 26 Jan 2021 16:51
 Operator : JC/MD
 Sample : M1218-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 237

Quant Time: Jan 27 03:30:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	210839	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	364509	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	313762	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	115485	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	137107	49.43	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.86%		
35) Dibromofluoromethane	7.550	113	107931	49.97	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.94%		
50) Toluene-d8	10.058	98	455278	53.62	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.24%		
62) 4-Bromofluorobenzene	12.373	95	148815	48.67	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.34%		
Target Compounds						Qvalue
11) Tert butyl alcohol	4.763	59	44777	86.18	ug/l #	86
16) Acetone	3.788	43	9761	9.04	ug/l	96
40) Benzene	8.004	78	5239	0.47	ug/l	100
52) Toluene	10.122	92	5602	0.84	ug/l	96
68) m/p-Xylenes	11.585	106	2397	0.54	ug/l	87
69) o-Xylene	11.920	106	1009	0.23	ug/l	76
84) 1,2,4-Trimethylbenzene	13.010	105	4272	0.50	ug/l	96
95) Naphthalene	15.100	128	2655	3.78	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
Data File : VN065608.D
Acq On : 26 Jan 2021 16:51
Operator : JC/MD
Sample : M1218-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
237

Quant Time: Jan 27 03:30:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

