

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065609.D
 Acq On : 26 Jan 2021 17:16
 Operator : JC/MD
 Sample : M1188-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-SWITCH-GEAR

Quant Time: Jan 27 03:30:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	207822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	359569	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	313915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	120747	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	133555	48.85	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.70%
35) Dibromofluoromethane	7.550	113	105269	49.40	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.80%
50) Toluene-d8	10.058	98	444958	53.12	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.24%
62) 4-Bromofluorobenzene	12.373	95	152300	50.50	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.00%
Target Compounds						
16) Acetone	3.785	43	29111	27.35	ug/l	98
20) Methylene Chloride	4.505	84	7486	2.54	ug/l #	92
43) Isopropyl Acetate	8.129	43	20474	3.33	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
Data File : VN065609.D
Acq On : 26 Jan 2021 17:16
Operator : JC/MD
Sample : M1188-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-SWITCH-GEAR

Quant Time: Jan 27 03:30:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

