

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065612.D
 Acq On : 26 Jan 2021 18:29
 Operator : JC/MD
 Sample : M1220-05
 Misc : 3.75g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP1-C

Quant Time: Jan 27 03:31:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	212131	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	365122	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	317332	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	114938	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	136005	48.73	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.46%
35) Dibromofluoromethane	7.550	113	108502	50.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.30%
50) Toluene-d8	10.058	98	460020	54.09	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.18%
62) 4-Bromofluorobenzene	12.373	95	150836	49.25	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
Data File : VN065612.D
Acq On : 26 Jan 2021 18:29
Operator : JC/MD
Sample : M1220-05
Misc : 3.75g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP1-C

Quant Time: Jan 27 03:31:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

