

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065613.D
 Acq On : 26 Jan 2021 18:54
 Operator : JC/MD
 Sample : M1218-01ME
 Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240ME

Quant Time: Jan 27 03:32:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	226977	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	394766	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	335927	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	129517	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	145774	48.82	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.64%
35) Dibromofluoromethane	7.547	113	114675	49.02	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.04%
50) Toluene-d8	10.055	98	486367	52.89	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.78%
62) 4-Bromofluorobenzene	12.373	95	160942	48.60	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.20%
Target Compounds						
52) Toluene	10.122	92	106386	14.80	ug/l	98
84) 1,2,4-Trimethylbenzene	13.010	105	13917	1.45	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
Data File : VN065613.D
Acq On : 26 Jan 2021 18:54
Operator : JC/MD
Sample : M1218-01ME
Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240ME

Quant Time: Jan 27 03:32:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

