

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065615.D
 Acq On : 26 Jan 2021 19:42
 Operator : JC/MD
 Sample : M1230-01MEDL 20X
 Misc : 1.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NWB-1501MEDL

Quant Time: Jan 27 03:32:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	203829	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	358643	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	313620	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	117534	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	133406	49.75	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.50%
35) Dibromofluoromethane	7.550	113	105508	49.64	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.28%
50) Toluene-d8	10.058	98	446723	53.47	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.94%
62) 4-Bromofluorobenzene	12.373	95	151265	50.28	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.56%
Target Compounds						
					Qvalue	
16) Acetone	3.785	43	20048	19.20	ug/l	100
52) Toluene	10.122	92	107581	16.48	ug/l	98
84) 1,2,4-Trimethylbenzene	13.010	105	14096	1.61	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
Data File : VN065615.D
Acq On : 26 Jan 2021 19:42
Operator : JC/MD
Sample : M1230-01MEDL 20X
Misc : 1.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NWB-1501MEDL

Quant Time: Jan 27 03:32:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

