

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065617.D
 Acq On : 26 Jan 2021 20:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 03:33:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.560	0.512	8.6	89	0.00
3 P	Chloromethane	0.823	0.749	9.0	89	0.00
4 C	Vinyl Chloride	0.833	0.767	7.9#	91	0.00
5 T	Bromomethane	0.482	0.468	2.9	95	0.00
6 T	Chloroethane	0.482	0.463	3.9	92	0.00
7 T	Trichlorofluoromethane	0.871	0.859	1.4	96	0.00
8 T	Diethyl Ether	0.415	0.412	0.7	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.529	0.523	1.1	99	0.00
10 T	Methyl Iodide	0.775	0.785	-1.3	99	0.00
11 T	Tert butyl alcohol	0.123	0.119	3.3	94	0.00
12 CM	1,1-Dichloroethene	0.609	0.576	5.4#	98	0.00
13 T	Acrolein	0.095	0.102	-7.4	96	0.00
14 T	Allyl chloride	1.128	1.132	-0.4	97	0.00
15 T	Acrylonitrile	0.321	0.326	-1.6	98	0.00
16 T	Acetone	0.256	0.238	7.0	93	0.00
17 T	Carbon Disulfide	1.916	1.705	11.0	95	0.00
18 T	Methyl Acetate	0.718	0.704	1.9	100	0.00
19 T	Methyl tert-butyl Ether	2.032	2.038	-0.3	100	0.00
20 T	Methylene Chloride	0.709	0.680	4.1	99	0.00
21 T	trans-1,2-Dichloroethene	0.647	0.630	2.6	100	0.00
22 T	Diisopropyl ether	2.307	2.311	-0.2	99	0.00
23 T	Vinyl Acetate	1.842	1.874	-1.7	98	0.00
24 P	1,1-Dichloroethane	1.229	1.249	-1.6	99	0.00
25 T	2-Butanone	0.402	0.408	-1.5	96	0.00
26 T	2,2-Dichloropropane	0.984	0.937	4.8	95	0.00
27 T	cis-1,2-Dichloroethene	0.741	0.737	0.5	100	0.00
28 T	Bromochloromethane	0.517	0.479	7.4	98	0.00
29 T	Tetrahydrofuran	0.291	0.277	4.8	96	0.00
30 C	Chloroform	1.132	1.148	-1.4#	101	0.00
31 T	Cyclohexane	1.237	1.123	9.2	94	0.00
32 T	1,1,1-Trichloroethane	0.922	0.926	-0.4	100	0.00
33 S	1,2-Dichloroethane-d4	0.658	0.635	3.5	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.296	0.281	5.1	101	0.00
36 T	1,1-Dichloropropene	0.503	0.494	1.8	100	0.00
37 T	Ethyl Acetate	0.483	0.492	-1.9	99	0.00
38 T	Carbon Tetrachloride	0.414	0.409	1.2	97	0.00
39 T	Methylcyclohexane	0.614	0.559	9.0	92	0.00
40 TM	Benzene	1.537	1.518	1.2	98	0.00
41 T	Methacrylonitrile	0.259	0.236	8.9	94	0.00
42 TM	1,2-Dichloroethane	0.463	0.483	-4.3	102	0.00
43 T	Isopropyl Acetate	0.856	0.824	3.7	98	0.00
44 TM	Trichloroethene	0.373	0.365	2.1	99	0.00
45 C	1,2-Dichloropropane	0.423	0.421	0.5#	99	0.00
46 T	Dibromomethane	0.234	0.236	-0.9	100	0.00
47 T	Bromodichloromethane	0.501	0.496	1.0	99	0.00
48 T	Methyl methacrylate	0.406	0.406	0.0	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065617.D
 Acq On : 26 Jan 2021 20:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 03:33:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	96	0.00
50 S	Toluene-d8	1.165	1.091	6.4	101	0.00
51 T	4-Methyl-2-Pentanone	0.472	0.466	1.3	97	0.00
52 CM	Toluene	0.910	0.911	-0.1#	101	0.00
53 T	t-1,3-Dichloropropene	0.583	0.581	0.3	99	0.00
54 T	cis-1,3-Dichloropropene	0.650	0.644	0.9	98	0.00
55 T	1,1,2-Trichloroethane	0.347	0.347	0.0	100	0.00
56 T	Ethyl methacrylate	0.574	0.579	-0.9	99	0.00
57 T	1,3-Dichloropropane	0.611	0.623	-2.0	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.265	0.256	3.4	96	0.00
59 T	2-Hexanone	0.328	0.332	-1.2	98	0.00
60 T	Dibromochloromethane	0.355	0.352	0.8	99	0.00
61 T	1,2-Dibromoethane	0.346	0.354	-2.3	101	0.00
62 S	4-Bromofluorobenzene	0.419	0.408	2.6	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.419	0.391	6.7	99	0.00
65 PM	Chlorobenzene	1.046	1.038	0.8	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.365	1.4	99	0.00
67 C	Ethyl Benzene	1.965	1.934	1.6#	100	0.00
68 T	m/p-Xylenes	0.711	0.705	0.8	100	0.00
69 T	o-Xylene	0.689	0.683	0.9	101	0.00
70 T	Styrene	1.135	1.156	-1.9	100	0.00
71 P	Bromoform	0.256	0.252	1.6	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	4.378	4.198	4.1	99	0.00
74 T	N-amyl acetate	1.852	2.188	-18.1	116	0.00
75 P	1,1,2,2-Tetrachloroethane	1.293	1.234	4.6	100	0.00
76 T	1,2,3-Trichloropropane	1.307	1.240	5.1	99	0.00
77 T	Bromobenzene	0.948	0.936	1.3	102	0.00
78 T	n-propylbenzene	5.159	4.952	4.0	99	0.00
79 T	2-Chlorotoluene	3.064	2.920	4.7	100	0.00
80 T	1,3,5-Trimethylbenzene	3.727	3.540	5.0	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.510	0.476	6.7	95	0.00
82 T	4-Chlorotoluene	3.128	2.981	4.7	99	0.00
83 T	tert-Butylbenzene	3.108	2.902	6.6	98	0.00
84 T	1,2,4-Trimethylbenzene	3.718	3.575	3.8	101	0.00
85 T	sec-Butylbenzene	4.196	3.927	6.4	98	0.00
86 T	p-Isopropyltoluene	3.705	3.567	3.7	100	0.00
87 T	1,3-Dichlorobenzene	1.690	1.664	1.5	100	0.00
88 T	1,4-Dichlorobenzene	1.715	1.665	2.9	102	0.00
89 T	n-Butylbenzene	3.329	3.206	3.7	98	0.00
90 T	Hexachloroethane	0.648	0.625	3.5	97	0.00
91 T	1,2-Dichlorobenzene	1.676	1.652	1.4	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.261	-1.6	100	0.00
93 T	1,2,4-Trichlorobenzene	0.891	0.955	-7.2	101	0.00
94 T	Hexachlorobutadiene	0.461	0.364	21.0	91	0.00
95 T	Naphthalene	2.867	3.252	-13.4	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
Data File : VN065617.D
Acq On : 26 Jan 2021 20:31
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 27 03:33:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.922	0.952	-3.3	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6