

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065599.D
 Acq On : 26 Jan 2021 13:11
 Operator : JC/MD
 Sample : M1244-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Jan 27 03:27:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	212165	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	367575	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	318637	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	109754	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	135305	48.48	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.96%
35) Dibromofluoromethane	7.547	113	108582	49.85	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.70%
50) Toluene-d8	10.055	98	456880	53.36	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.72%
62) 4-Bromofluorobenzene	12.373	95	151259	49.06	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.12%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
Data File : VN065599.D
Acq On : 26 Jan 2021 13:11
Operator : JC/MD
Sample : M1244-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Jan 27 03:27:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.621 min Scan# 1829
Delta R.T. -0.003 min
Lab File: VN065599.D
Acq: 26 Jan 2021 13:11




Instrument :
MSVOA_N
ClientSampled :
TB


Tgt Ion:168 Resp: 212165
Ion Ratio Lower Upper
168 100
99 61.0 42.6 63.8



#33
1,2-Dichloroethane-d4
Concen: 48.48 ug/l
RT: 7.984 min Scan# 1942
Delta R.T. -0.003 min
Lab File: VN065599.D
Acq: 26 Jan 2021 13:11

Tgt Ion: 65 Resp: 135305
Ion Ratio Lower Upper
65 100
67 55.5 0.0 107.2









#34

1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.547 min Scan# 2117
Delta R.T. -0.003 min
Lab File: VN065599.D
Acq: 26 Jan 2021 13:11

Instrument :
MSVOA_N
ClientSampled :
TB




Tgt Ion:114 Resp: 367575
Ion Ratio Lower Upper
114 100
63 22.6 0.0 38.6
88 16.7 0.0 30.6




#35

Dibromofluoromethane
Concen: 49.85 ug/l
RT: 7.547 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN065599.D
Acq: 26 Jan 2021 13:11

Tgt Ion:113 Resp: 108582
Ion Ratio Lower Upper
113 100
111 102.2 82.2 123.2
192 17.5 15.4 23.0









#50
Toluene-d8
Concen: 53.36 ug/l
RT: 10.055 min Scan# 2586
Delta R.T. -0.003 min
Lab File: VN065599.D
Acq: 26 Jan 2021 13:11




Instrument :
MSVOA_N
ClientSampled :
TB


Tgt Ion: 98 Resp: 456880
Ion Ratio Lower Upper
98 100
100 64.6 50.5 75.7



#62
4-Bromofluorobenzene
Concen: 49.06 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. -0.000 min
Lab File: VN065599.D
Acq: 26 Jan 2021 13:11

Tgt Ion: 95 Resp: 151259
Ion Ratio Lower Upper
95 100
174 66.9 0.0 169.0
176 65.0 0.0 162.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.376 min Scan# 2997
Delta R.T. 0.000 min
Lab File: VN065599.D
Acq: 26 Jan 2021 13:11

Instrument :
MSVOA_N
ClientSampled :
TB

Tgt Ion:117 Resp: 318637
Ion Ratio Lower Upper
117 100
82 59.0 43.7 65.5
119 31.9 26.1 39.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.315 min Scan# 3600
Delta R.T. 0.003 min
Lab File: VN065599.D
Acq: 26 Jan 2021 13:11

Tgt Ion:152 Resp: 109754
Ion Ratio Lower Upper
152 100
115 62.8 28.5 85.6
150 160.6 0.0 347.6

