


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012621\
 Data File : VN065592.D
 Acq On : 26 Jan 2021 8:48
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Title : SW846 8260
 Last Update : Mon Jan 25 13:47:28 2021



AutoFind: Scans 3306, 3307, 3308; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	4282	PASS
75	95	30	60	45.8	11343	PASS
95	95	100	100	100.0	24747	PASS
96	95	5	9	6.4	1573	PASS
173	174	0.00	2	1.1	182	PASS
174	95	50	100	67.0	16577	PASS
175	174	5	9	8.2	1365	PASS
176	174	95	101	95.1	15770	PASS
177	176	5	9	6.6	1046	PASS