

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
 Data File : VN070783.D
 Acq On : 26 Jan 2022 12:12
 Operator : JC/MD
 Sample : VN0126MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0126MBL01

Quant Time: Jan 27 03:31:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	865283	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.965	114	1527294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1188995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	409471	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	636651	51.339	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.680%
35) Dibromofluoromethane	8.021	113	405210	42.772	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	85.540%
50) Toluene-d8	10.440	98	1957203	49.039	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.728	95	522875	36.765	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	73.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
Data File : VN070783.D
Acq On : 26 Jan 2022 12:12
Operator : JC/MD
Sample : VN0126MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0126MBL01

Quant Time: Jan 27 03:31:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

