

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
 Data File : VN070784.D
 Acq On : 26 Jan 2022 12:38
 Operator : JC/MD
 Sample : VN0126WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0126WBL01

Quant Time: Jan 27 03:31:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	815321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1296105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1236767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	406054	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	566662	48.496	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.000%
35) Dibromofluoromethane	8.021	113	385720	47.977	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.960%
50) Toluene-d8	10.441	98	1730933	51.106	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.220%
62) 4-Bromofluorobenzene	12.728	95	561676	46.538	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
Data File : VN070784.D
Acq On : 26 Jan 2022 12:38
Operator : JC/MD
Sample : VN0126WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0126WBL01

Quant Time: Jan 27 03:31:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

