

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
 Data File : VN070793.D
 Acq On : 26 Jan 2022 16:39
 Operator : JC/MD
 Sample : N1267-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BASMENT-FLOORS

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/27/2022
 Supervised By : Mahesh Dadoda 01/27/2022

Quant Time: Jan 27 03:33:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	844277	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1355218	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1224678	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	571201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	593251	49.030	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.060%		
35) Dibromofluoromethane	8.021	113	403652	48.017	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.040%		
50) Toluene-d8	10.440	98	1711525	48.329	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.660%		
62) 4-Bromofluorobenzene	12.731	95	715584	56.704	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	113.400%		
Target Compounds						
20) Methylene Chloride	5.103	84	20118m	1.819	ug/l	
43) Isopropyl Acetate	8.560	43	89787	3.129	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
Data File : VN070793.D
Acq On : 26 Jan 2022 16:39
Operator : JC/MD
Sample : N1267-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BASMENT-FLOORS

Quant Time: Jan 27 03:33:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/27/2022
Supervised By :Mahesh Dadoda 01/27/2022

