

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
 Data File : VN070795.D
 Acq On : 26 Jan 2022 17:30
 Operator : JC/MD
 Sample : N1267-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2ND-3RD-FLOOR-WALLS-1

Quant Time: Jan 27 03:33:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	820845	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1304855	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1167749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	430599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	571743	48.601	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.200%
35) Dibromofluoromethane	8.021	113	389422	48.112	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.220%
50) Toluene-d8	10.440	98	1646611	48.290	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.580%
62) 4-Bromofluorobenzene	12.731	95	529841	43.606	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.220%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	61569	7.541	ug/l	95
20) Methylene Chloride	5.098	84	26561	2.470	ug/l	91
43) Isopropyl Acetate	8.560	43	100005	3.619	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
Data File : VN070795.D
Acq On : 26 Jan 2022 17:30
Operator : JC/MD
Sample : N1267-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2ND-3RD-FLOOR-WALLS-1

Quant Time: Jan 27 03:33:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

