

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
 Data File : VN070796.D
 Acq On : 26 Jan 2022 17:55
 Operator : JC/MD
 Sample : N1267-23
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2ND-3RD-FLOOR-WALLS-2

Quant Time: Jan 27 03:33:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/27/2022
 Supervised By : Mahesh Dadoda 01/27/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	814889	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	1310900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1167635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	429943	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	560392	47.984	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.960%		
35) Dibromofluoromethane	8.021	113	382751	47.070	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.140%		
50) Toluene-d8	10.440	98	1634601	47.717	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.440%		
62) 4-Bromofluorobenzene	12.731	95	518186	42.450	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	84.900%		
Target Compounds						
						Qvalue
16) Acetone	4.296	43	43194	5.329	ug/l	98
20) Methylene Chloride	5.103	84	18488m	1.732	ug/l	
43) Isopropyl Acetate	8.560	43	87530	3.153	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012622\
Data File : VN070796.D
Acq On : 26 Jan 2022 17:55
Operator : JC/MD
Sample : N1267-23
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2ND-3RD-FLOOR-WALLS-2

Quant Time: Jan 27 03:33:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/27/2022
Supervised By :Mahesh Dadoda 01/27/2022

