

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
 Data File : VN076416.D
 Acq On : 26 Jan 2023 13:37
 Operator : JC\MD
 Sample : PB150465ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150465ZHE#01

Quant Time: Jan 27 00:34:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	474973	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	789383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	666245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	260534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	297407	55.922	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.840%
35) Dibromofluoromethane	8.177	113	245832	49.785	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.560%
50) Toluene-d8	10.570	98	879718	48.240	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.480%
62) 4-Bromofluorobenzene	12.853	95	276095	46.377	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.760%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	37793	22.101	ug/l	# 87
20) Methylene Chloride	5.294	84	33434	6.099	ug/l	97
37) Ethyl Acetate	7.571	43	41479	7.401	ug/l	99
43) Isopropyl Acetate	8.694	43	353381	38.597	ug/l	# 74
95) Naphthalene	15.647	128	5121	2.203	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
Data File : VN076416.D
Acq On : 26 Jan 2023 13:37
Operator : JC\MD
Sample : PB150465ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150465ZHE#01

Quant Time: Jan 27 00:34:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

