

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
 Data File : VN076418.D
 Acq On : 26 Jan 2023 14:25
 Operator : JC\MD
 Sample : PB150465ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150465ZHE#03

Quant Time: Jan 27 00:35:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	397125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	665102	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	592259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	228550	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	233126	52.429	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.860%		
35) Dibromofluoromethane	8.171	113	201926	48.534	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.060%		
50) Toluene-d8	10.571	98	778782	50.685	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.360%		
62) 4-Bromofluorobenzene	12.853	95	247348	49.312	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.620%		
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	38104	26.666	ug/l	91
20) Methylene Chloride	5.283	84	29870	6.558	ug/l	96
37) Ethyl Acetate	7.571	43	35723	7.565	ug/l #	95
43) Isopropyl Acetate	8.694	43	416515	53.993	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
Data File : VN076418.D
Acq On : 26 Jan 2023 14:25
Operator : JC\MD
Sample : PB150465ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150465ZHE#03

Quant Time: Jan 27 00:35:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

