

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
 Data File : VN076421.D
 Acq On : 26 Jan 2023 15:37
 Operator : JC\MD
 Sample : 01270-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-1-012323

Quant Time: Jan 27 00:37:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	383737	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	639645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	579824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	230947	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	232946	54.216	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.440%		
35) Dibromofluoromethane	8.171	113	198184	49.531	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.060%		
50) Toluene-d8	10.571	98	762904	51.628	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.260%		
62) 4-Bromofluorobenzene	12.853	95	243948	50.569	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 101.140%		
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	37141	26.900	ug/l	92
20) Methylene Chloride	5.300	84	15790	3.312	ug/l #	90
37) Ethyl Acetate	7.565	43	37915	8.344	ug/l	96
43) Isopropyl Acetate	8.694	43	264687	35.677	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
Data File : VN076421.D
Acq On : 26 Jan 2023 15:37
Operator : JC\MD
Sample : 01270-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-1-012323

Quant Time: Jan 27 00:37:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

