

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
 Data File : VN076422.D
 Acq On : 26 Jan 2023 16:01
 Operator : JC\MD
 Sample : 01283-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-H

Quant Time: Jan 27 00:37:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	442443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	746370	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	653884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	258176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	266597	53.815	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.620%
35) Dibromofluoromethane	8.177	113	224613	48.109	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.220%
50) Toluene-d8	10.571	98	875725	50.788	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.580%
62) 4-Bromofluorobenzene	12.853	95	272020	48.325	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.660%
Target Compounds						
						Qvalue
16) Acetone	4.453	43	39177	24.603	ug/l	98
20) Methylene Chloride	5.289	84	15771	2.787	ug/l	91
37) Ethyl Acetate	7.571	43	36143	6.824	ug/l	98
43) Isopropyl Acetate	8.694	43	270582	31.257	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
Data File : VN076422.D
Acq On : 26 Jan 2023 16:01
Operator : JC\MD
Sample : 01283-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-H

Quant Time: Jan 27 00:37:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

