

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
 Data File : VN076423.D
 Acq On : 26 Jan 2023 16:25
 Operator : JC\MD
 Sample : 01283-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-10

Quant Time: Jan 27 00:38:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	386263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	650755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	596919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	236260	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	234471	54.214	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.420%	
35) Dibromofluoromethane	8.171	113	197591	48.539	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.080%	
50) Toluene-d8	10.571	98	778961	51.814	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.620%	
62) 4-Bromofluorobenzene	12.853	95	254536	51.863	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.720%	
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	37944	27.302	ug/l	97
20) Methylene Chloride	5.289	84	16106	3.364	ug/l	97
37) Ethyl Acetate	7.571	43	37019	8.010	ug/l	98
43) Isopropyl Acetate	8.694	43	330841	43.833	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
Data File : VN076423.D
Acq On : 26 Jan 2023 16:25
Operator : JC\MD
Sample : 01283-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-10

Quant Time: Jan 27 00:38:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

