

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
 Data File : VN076425.D
 Acq On : 26 Jan 2023 17:13
 Operator : JC\MD
 Sample : 01289-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ209-STOCK

Quant Time: Jan 27 00:39:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	374063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	620822	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	556121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	222924	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	225050	53.733	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.460%	
35) Dibromofluoromethane	8.171	113	189471	48.789	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.580%	
50) Toluene-d8	10.571	98	732777	51.092	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.180%	
62) 4-Bromofluorobenzene	12.847	95	233228	49.813	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.620%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	45160	33.571	ug/l	98
20) Methylene Chloride	5.289	84	12590	2.598	ug/l	94
25) 2-Butanone	7.494	43	12830	6.257	ug/l	93
37) Ethyl Acetate	7.565	43	36823	8.350	ug/l #	91
43) Isopropyl Acetate	8.694	43	270758	37.602	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
Data File : VN076425.D
Acq On : 26 Jan 2023 17:13
Operator : JC\MD
Sample : 01289-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ209-STOCK

Quant Time: Jan 27 00:39:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

