

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
 Data File : VN076410.D
 Acq On : 26 Jan 2023 09:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 00:29:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	0.512	0.486	5.1	102	0.00
3 P	Chloromethane	0.531	0.516	2.8	110	0.00
4 C	Vinyl Chloride	0.650	0.720	-10.8#	124	0.00
5 T	Bromomethane	0.533	0.527	1.1	113	0.00
6 T	Chloroethane	0.488	0.514	-5.3	125	0.00
7 T	Trichlorofluoromethane	0.895	0.963	-7.6	122	0.00
8 T	Diethyl Ether	0.300	0.333	-11.0	120	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.507	0.525	-3.6	117	0.00
10 T	Methyl Iodide	0.774	0.736	4.9	104	0.00
11 T	Tert butyl alcohol	0.075	0.070	6.7	103	0.01
12 CM	1,1-Dichloroethene	0.479	0.467	2.5#	108	0.00
13 T	Acrolein	0.090	0.114	-26.7#	144	0.00
14 T	Allyl chloride	0.558	0.577	-3.4	116	0.00
15 T	Acrylonitrile	0.213	0.221	-3.8	114	0.00
16 T	Acetone	0.195	0.255	-30.8#	161#	0.00
17 T	Carbon Disulfide	1.192	1.039	12.8	97	0.00
18 T	Methyl Acetate	0.608	0.626	-3.0	118	0.00
19 T	Methyl tert-butyl Ether	1.559	1.599	-2.6	111	0.00
20 T	Methylene Chloride	0.645	0.526	18.4	113	0.00
21 T	trans-1,2-Dichloroethene	0.519	0.484	6.7	106	0.00
22 T	Diisopropyl ether	1.316	1.448	-10.0	119	0.00
23 T	Vinyl Acetate	0.851	0.940	-10.5	114	0.00
24 P	1,1-Dichloroethane	0.877	0.878	-0.1	112	0.00
25 T	2-Butanone	0.274	0.303	-10.6	123	0.00
26 T	2,2-Dichloropropane	0.739	0.766	-3.7	110	0.00
27 T	cis-1,2-Dichloroethene	0.605	0.596	1.5	110	0.00
28 T	Bromochloromethane	0.356	0.412	-15.7	126	0.00
29 T	Tetrahydrofuran	0.169	0.182	-7.7	114	0.00
30 C	Chloroform	0.971	1.011	-4.1#	116	0.00
31 T	Cyclohexane	0.860	0.754	12.3	106	0.00
32 T	1,1,1-Trichloroethane	0.886	0.906	-2.3	114	0.00
33 S	1,2-Dichloroethane-d4	0.560	0.635	-13.4	127	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	114	0.00
35 S	Dibromofluoromethane	0.313	0.332	-6.1	123	0.00
36 T	1,1-Dichloropropene	0.444	0.431	2.9	110	0.00
37 T	Ethyl Acetate	0.391	0.361	7.7	117	0.00
38 T	Carbon Tetrachloride	0.496	0.484	2.4	110	0.00
39 T	Methylcyclohexane	0.543	0.533	1.8	106	0.00
40 TM	Benzene	1.329	1.290	2.9	110	0.00
41 T	Methacrylonitrile	0.187	0.197	-5.3	117	0.00
42 TM	1,2-Dichloroethane	0.455	0.473	-4.0	120	0.00
43 T	Isopropyl Acetate	0.580	0.606	-4.5	115	0.00
44 TM	Trichloroethene	0.376	0.347	7.7	109	0.00
45 C	1,2-Dichloropropane	0.318	0.324	-1.9#	117	0.00
46 T	Dibromomethane	0.240	0.241	-0.4	114	0.00
47 T	Bromodichloromethane	0.466	0.498	-6.9	119	0.00
48 T	Methyl methacrylate	0.275	0.290	-5.5	117	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
 Data File : VN076410.D
 Acq On : 26 Jan 2023 09:58
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 00:29:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	105	0.00
50 S	Toluene-d8	1.155	1.203	-4.2	117	0.00
51 T	4-Methyl-2-Pentanone	0.340	0.366	-7.6	117	0.00
52 CM	Toluene	0.889	0.860	3.3#	110	0.00
53 T	t-1,3-Dichloropropene	0.444	0.494	-11.3	115	0.00
54 T	cis-1,3-Dichloropropene	0.511	0.544	-6.5	116	0.00
55 T	1,1,2-Trichloroethane	0.340	0.344	-1.2	118	0.00
56 T	Ethyl methacrylate	0.439	0.494	-12.5	117	0.00
57 T	1,3-Dichloropropane	0.552	0.567	-2.7	116	0.00
58 T	2-Chloroethyl Vinyl ether	0.148	0.144	2.7	102	0.00
59 T	2-Hexanone	0.247	0.274	-10.9	119	0.00
60 T	Dibromochloromethane	0.358	0.392	-9.5	117	0.00
61 T	1,2-Dibromoethane	0.339	0.350	-3.2	114	0.00
62 S	4-Bromofluorobenzene	0.377	0.433	-14.9	125	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	114	0.00
64 T	Tetrachloroethene	0.451	0.363	19.5	105	0.00
65 PM	Chlorobenzene	1.057	1.011	4.4	112	0.00
66 T	1,1,1,2-Tetrachloroethane	0.393	0.394	-0.3	115	0.00
67 C	Ethyl Benzene	1.809	1.841	-1.8#	112	0.00
68 T	m/p-Xylenes	0.725	0.725	0.0	111	0.00
69 T	o-Xylene	0.712	0.718	-0.8	111	0.00
70 T	Styrene	1.137	1.183	-4.0	112	0.00
71 P	Bromoform	0.284	0.303	-6.7	114	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
73 T	Isopropylbenzene	3.868	3.808	1.6	112	0.00
74 T	N-amyl acetate	1.113	1.181	-6.1	116	0.00
75 P	1,1,2,2-Tetrachloroethane	1.112	1.063	4.4	113	0.00
76 T	1,2,3-Trichloropropane	1.034	0.963	6.9	118	0.00
77 T	Bromobenzene	0.994	0.877	11.8	110	0.00
78 T	n-propylbenzene	4.364	4.391	-0.6	113	0.00
79 T	2-Chlorotoluene	2.678	2.590	3.3	114	0.00
80 T	1,3,5-Trimethylbenzene	3.240	3.239	0.0	112	0.00
81 T	trans-1,4-Dichloro-2-butene	0.301	0.315	-4.7	110	0.00
82 T	4-Chlorotoluene	2.678	2.590	3.3	114	0.00
83 T	tert-Butylbenzene	2.882	2.918	-1.2	115	0.00
84 T	1,2,4-Trimethylbenzene	3.249	3.244	0.2	112	0.00
85 T	sec-Butylbenzene	3.999	4.049	-1.3	112	0.00
86 T	p-Isopropyltoluene	3.289	3.451	-4.9	112	0.00
87 T	1,3-Dichlorobenzene	1.823	1.700	6.7	111	0.00
88 T	1,4-Dichlorobenzene	1.842	1.650	10.4	109	0.00
89 T	n-Butylbenzene	2.611	2.844	-8.9	113	0.00
90 T	Hexachloroethane	0.568	0.586	-3.2	112	0.00
91 T	1,2-Dichlorobenzene	1.816	1.653	9.0	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.193	0.196	-1.6	106	0.00
93 T	1,2,4-Trichlorobenzene	0.870	0.834	4.1	107	0.00
94 T	Hexachlorobutadiene	0.482	0.432	10.4	102	0.00
95 T	Naphthalene	2.543	2.416	5.0	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012623\
Data File : VN076410.D
Acq On : 26 Jan 2023 09:58
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 27 00:29:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.823	0.804	2.3	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6