

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
 Data File : VN080765.D
 Acq On : 26 Jan 2024 10:53
 Operator : JC\MD
 Sample : VN0126MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0126MBL01

Quant Time: Jan 29 02:27:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	112706	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	254635	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	260475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103687	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	85419	43.158	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.320%
35) Dibromofluoromethane	8.165	113	71568	49.467	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.940%
50) Toluene-d8	10.570	98	286906	52.007	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.020%
62) 4-Bromofluorobenzene	12.853	95	112540	53.198	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
Data File : VN080765.D
Acq On : 26 Jan 2024 10:53
Operator : JC\MD
Sample : VN0126MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0126MBL01

Quant Time: Jan 29 02:27:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

