

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
 Data File : VN080772.D
 Acq On : 26 Jan 2024 16:12
 Operator : JC\MD
 Sample : P1240-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11985

Quant Time: Jan 29 02:32:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	114481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	266272	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	288303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	90912	45.221	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.440%
35) Dibromofluoromethane	8.165	113	73676	48.699	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	10.571	98	309767	53.697	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.400%
62) 4-Bromofluorobenzene	12.853	95	122946	55.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.160%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	21735	32.930	ug/l	97
20) Methylene Chloride	5.283	84	9091	4.718	ug/l #	65
25) 2-Butanone	7.494	43	15480	12.975	ug/l #	64
37) Ethyl Acetate	7.571	43	4558	1.517	ug/l #	72
43) Isopropyl Acetate	8.688	43	140013	25.347	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
Data File : VN080772.D
Acq On : 26 Jan 2024 16:12
Operator : JC\MD
Sample : P1240-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11985

Quant Time: Jan 29 02:32:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

