

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
 Data File : VN080774.D
 Acq On : 26 Jan 2024 17:00
 Operator : JC\MD
 Sample : P1240-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-223

Quant Time: Jan 29 02:33:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	102693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	242094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	258102	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	105344	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	83251	46.164	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.320%
35) Dibromofluoromethane	8.171	113	65667	47.740	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.480%
50) Toluene-d8	10.571	98	281272	53.627	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.260%
62) 4-Bromofluorobenzene	12.853	95	112981	56.173	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.340%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.512	59	2500	8.197	ug/l #	79
16) Acetone	4.442	43	22088	37.307	ug/l	97
20) Methylene Chloride	5.271	84	8961	5.185	ug/l #	79
25) 2-Butanone	7.494	43	17245	16.114	ug/l #	75
43) Isopropyl Acetate	8.694	43	135062	26.892	ug/l #	81
95) Naphthalene	15.641	128	105169	14.328	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
Data File : VN080774.D
Acq On : 26 Jan 2024 17:00
Operator : JC\MD
Sample : P1240-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-223

Quant Time: Jan 29 02:33:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

