

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
 Data File : VN080775.D
 Acq On : 26 Jan 2024 17:24
 Operator : JC\MD
 Sample : P1243-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-308-0-2

Quant Time: Jan 29 02:34:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	104669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	240789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	254314	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97438	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	83972	45.684	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.360%
35) Dibromofluoromethane	8.171	113	64875	47.420	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.840%
50) Toluene-d8	10.571	98	272729	52.280	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.560%
62) 4-Bromofluorobenzene	12.853	95	106628	53.301	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.600%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	23307	38.622	ug/l #	87
20) Methylene Chloride	5.271	84	5555	3.153	ug/l #	77
43) Isopropyl Acetate	8.694	43	130314	26.088	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
Data File : VN080775.D
Acq On : 26 Jan 2024 17:24
Operator : JC\MD
Sample : P1243-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-308-0-2

Quant Time: Jan 29 02:34:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

