

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
 Data File : VN080777.D
 Acq On : 26 Jan 2024 18:12
 Operator : JC\MD
 Sample : P1246-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-COMP-012424

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/29/2024
 Supervised By : Mahesh Dadoda 01/29/2024

Quant Time: Jan 29 02:35:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	100347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	232074	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	255815	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	83410	47.333	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.660%		
35) Dibromofluoromethane	8.165	113	66661	50.555	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.100%		
50) Toluene-d8	10.571	98	281676	56.023	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	112.040%		
62) 4-Bromofluorobenzene	12.847	95	113945	59.098	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	118.200%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	21287	36.794	ug/l	94
20) Methylene Chloride	5.277	84	8647m	5.120	ug/l	
25) 2-Butanone	7.483	43	15051	14.392	ug/l	# 63
43) Isopropyl Acetate	8.688	43	143027	29.708	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012624\
Data File : VN080777.D
Acq On : 26 Jan 2024 18:12
Operator : JC\MD
Sample : P1246-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-COMP-012424

Quant Time: Jan 29 02:35:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/29/2024
Supervised By :Mahesh Dadoda 01/29/2024

