

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012720\
 Data File : VN059873.D
 Acq On : 27 Jan 2020 15:29
 Operator : JC/MD
 Sample : L1010-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-01-012420

Quant Time: Jan 27 23:16:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	219782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	330860	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	288947	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	122196	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	115275	47.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.46%	
35) Dibromofluoromethane	7.57	113	98453	49.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.80%	
50) Toluene-d8	10.08	98	386177	50.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.04%	
62) 4-Bromofluorobenzene	12.40	95	128438	46.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.22%	
Target Compounds						
20) Methylene Chloride	4.52	84	22112	8.951	ug/l	90
43) Isopropyl Acetate	8.14	43	41641	8.959	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012720\
Data File : VN059873.D
Acq On : 27 Jan 2020 15:29
Operator : JC/MD
Sample : L1010-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-01-012420

Quant Time: Jan 27 23:16:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

