

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012721\
 Data File : VN065635.D
 Acq On : 27 Jan 2021 17:26
 Operator : JC/MD
 Sample : M1213-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IDW-BR2-SO-012121

Quant Time: Jan 28 02:17:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	206932	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	358615	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	318822	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	122725	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	131872	48.44	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.88%
35) Dibromofluoromethane	7.547	113	105738	49.76	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.52%
50) Toluene-d8	10.058	98	447314	53.55	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.10%
62) 4-Bromofluorobenzene	12.373	95	151878	50.49	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.98%
Target Compounds						
16) Acetone	3.782	43	27131	25.60	ug/l	93
20) Methylene Chloride	4.499	84	8520	2.90	ug/l	98
43) Isopropyl Acetate	8.132	43	20293	3.31	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012721\
Data File : VN065635.D
Acq On : 27 Jan 2021 17:26
Operator : JC/MD
Sample : M1213-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IDW-BR2-SO-012121

Quant Time: Jan 28 02:17:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

