

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012722\
 Data File : VN070801.D
 Acq On : 27 Jan 2022 12:52
 Operator : JC/MD
 Sample : VN0127WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0127WBL01

Quant Time: Jan 28 00:36:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1009757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1611256	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.741	117	1573344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	548298	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	710184	49.075	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.160%
35) Dibromofluoromethane	8.019	113	472454	47.271	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.540%
50) Toluene-d8	10.438	98	2201059	52.276	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.560%
62) 4-Bromofluorobenzene	12.728	95	730882	48.713	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012722\
Data File : VN070801.D
Acq On : 27 Jan 2022 12:52
Operator : JC/MD
Sample : VN0127WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0127WBL01

Quant Time: Jan 28 00:36:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

