

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012722\
 Data File : VN070807.D
 Acq On : 27 Jan 2022 15:40
 Operator : JC/MD
 Sample : N1278-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/28/2022
 Supervised By : Mahesh Dadoda 01/28/2022

Quant Time: Jan 28 00:37:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	869277	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.965	114	1389074	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1244956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	451514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	602507	48.363	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.720%		
35) Dibromofluoromethane	8.021	113	411452	47.752	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.500%		
50) Toluene-d8	10.441	98	1703838	46.939	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	93.880%		
62) 4-Bromofluorobenzene	12.731	95	573586	44.344	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.680%		
Target Compounds						
16) Acetone	4.293	43	11478m	1.327	ug/l	Qvalue
64) Tetrachloroethene	10.974	164	3878	0.425	ug/l	91
95) Naphthalene	15.507	128	4556	5.124	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012722\
Data File : VN070807.D
Acq On : 27 Jan 2022 15:40
Operator : JC/MD
Sample : N1278-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Jan 28 00:37:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/28/2022
Supervised By :Mahesh Dadoda 01/28/2022

