

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012722\
 Data File : VN070812.D
 Acq On : 27 Jan 2022 17:46
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 28 00:38:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	807101	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	1322339	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1180460	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	462791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	574306	49.650	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.300%
35) Dibromofluoromethane	8.021	113	387101	47.193	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.380%
50) Toluene-d8	10.440	98	1628020	47.114	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.220%
62) 4-Bromofluorobenzene	12.731	95	615340	49.973	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.940%
Target Compounds						
					Qvalue	
30) Chloroform	7.823	83	1769	Below Cal	#	66
95) Naphthalene	15.507	128	36598	6.323	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012722\
Data File : VN070812.D
Acq On : 27 Jan 2022 17:46
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 28 00:38:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

