

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012722\
 Data File : VN070815.D
 Acq On : 27 Jan 2022 19:02
 Operator : JC/MD
 Sample : N1287-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WASTE-CHARACTER-2

Quant Time: Jan 28 00:38:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	722140	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	1186767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1057315	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	393114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	512304	49.501	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.000%
35) Dibromofluoromethane	8.024	113	348558	47.349	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.700%
50) Toluene-d8	10.440	98	1463957	47.206	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.420%
62) 4-Bromofluorobenzene	12.731	95	475846	43.059	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.120%
Target Compounds						
					Qvalue	
16) Acetone	4.304	43	9731	1.355	ug/l	98
30) Chloroform	7.817	83	223773	12.902	ug/l	99
47) Bromodichloromethane	9.759	83	57480	4.629	ug/l	98
60) Dibromochloromethane	11.237	129	8525	1.013	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012722\
Data File : VN070815.D
Acq On : 27 Jan 2022 19:02
Operator : JC/MD
Sample : N1287-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WASTE-CHARACTER-2

Quant Time: Jan 28 00:38:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

