

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
 Data File : VN076429.D
 Acq On : 27 Jan 2023 10:19
 Operator : JC\MD
 Sample : VN0127MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0127MBL01

Quant Time: Jan 27 22:33:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	333797	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	539757	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	466480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	178867	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	198172	53.023	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.040%
35) Dibromofluoromethane	8.171	113	168025	49.764	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.520%
50) Toluene-d8	10.571	98	627450	50.319	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.640%
62) 4-Bromofluorobenzene	12.853	95	189795	46.625	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
Data File : VN076429.D
Acq On : 27 Jan 2023 10:19
Operator : JC\MD
Sample : VN0127MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0127MBL01

Quant Time: Jan 27 22:33:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

