

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
 Data File : VN076430.D
 Acq On : 27 Jan 2023 10:42
 Operator : JC\MD
 Sample : VN0127WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0127WBL01

Quant Time: Jan 27 22:33:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	387881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	636588	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	542840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	215925	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	229992	52.956	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.920%
35) Dibromofluoromethane	8.171	113	195829	49.177	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.360%
50) Toluene-d8	10.571	98	728218	49.517	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.040%
62) 4-Bromofluorobenzene	12.853	95	222664	46.379	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
Data File : VN076430.D
Acq On : 27 Jan 2023 10:42
Operator : JC\MD
Sample : VN0127WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0127WBL01

Quant Time: Jan 27 22:33:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

