

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
 Data File : VN076433.D
 Acq On : 27 Jan 2023 12:35
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 27 23:22:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	419061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	694898	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	596127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	236023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	255409	54.433	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.860%
35) Dibromofluoromethane	8.177	113	216790	49.873	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.740%
50) Toluene-d8	10.571	98	787902	49.080	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.160%
62) 4-Bromofluorobenzene	12.853	95	252371	48.156	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.320%
Target Compounds						
						Qvalue
93) 1,2,4-Trichlorobenzene	15.400	180	2084	1.795	ug/l	96
95) Naphthalene	15.641	128	9472	2.583	ug/l	97
96) 1,2,3-Trichlorobenzene	15.847	180	2555	1.786	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
Data File : VN076433.D
Acq On : 27 Jan 2023 12:35
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 27 23:22:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

