

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
 Data File : VN076437.D
 Acq On : 27 Jan 2023 14:11
 Operator : JC\MD
 Sample : 01332-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jan 27 22:37:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	395853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	645983	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	555093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	204005	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	237738	53.638	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.280%
35) Dibromofluoromethane	8.177	113	201794	49.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.880%
50) Toluene-d8	10.571	98	743785	49.840	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.680%
62) 4-Bromofluorobenzene	12.853	95	223071	45.788	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
Data File : VN076437.D
Acq On : 27 Jan 2023 14:11
Operator : JC\MD
Sample : 01332-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jan 27 22:37:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

