

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
 Data File : VN076441.D
 Acq On : 27 Jan 2023 15:58
 Operator : JC\MD
 Sample : 01329-10
 Misc : 7.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOTAL-VOC-7-BORING-2-14-21-GRAB-8

Quant Time: Jan 28 01:08:42 2023
 Quant Title :
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	350933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	571023	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	478560	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	202986	51.659	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.320%
35) Dibromofluoromethane	8.165	113	168385	47.140	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	10.565	98	660973	50.105	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.200%
62) 4-Bromofluorobenzene	12.853	95	200158	46.478	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.960%
Target Compounds						
49) 1,4-Dioxane	9.712	88	1970	29.776	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
Data File : VN076441.D
Acq On : 27 Jan 2023 15:58
Operator : JC\MD
Sample : 01329-10
Misc : 7.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOTAL-VOC-7-BORING-2-14-21-GRAB-8

Quant Time: Jan 28 01:08:42 2023
Quant Title :
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

