

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
 Data File : VN076443.D
 Acq On : 27 Jan 2023 16:46
 Operator : JC\MD
 Sample : 01329-20
 Misc : 16.75g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOTAL-VOC-BORING-2-25-35-GRAB-7

Quant Time: Jan 27 22:41:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	391049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	632542	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	525870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	209015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	225254	51.445	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.900%
35) Dibromofluoromethane	8.165	113	186168	47.050	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.100%
50) Toluene-d8	10.570	98	716598	49.038	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.853	95	214670	45.000	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.000%
Target Compounds						
						Qvalue
49) 1,4-Dioxane	9.706	88	1571	21.436	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012723\
Data File : VN076443.D
Acq On : 27 Jan 2023 16:46
Operator : JC\MD
Sample : 01329-20
Misc : 16.75g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOTAL-VOC-BORING-2-25-35-GRAB-7

Quant Time: Jan 27 22:41:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

