

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
 Data File : VN085518.D
 Acq On : 27 Jan 2025 15:48
 Operator : JC\MD
 Sample : VN0127WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0127WBL01

Quant Time: Jan 28 02:00:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	197682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	347425	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	116627	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	163263	51.164	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.320%		
35) Dibromofluoromethane	8.165	113	122921	50.999	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.000%		
50) Toluene-d8	10.565	98	427942	49.972	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.940%		
62) 4-Bromofluorobenzene	12.847	95	134001	45.743	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 91.480%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
Data File : VN085518.D
Acq On : 27 Jan 2025 15:48
Operator : JC\MD
Sample : VN0127WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0127WBL01

Quant Time: Jan 28 02:00:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

