

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
 Data File : VN085521.D
 Acq On : 27 Jan 2025 17:10
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 28 02:02:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	194988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	348925	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	162025	51.477	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.960%	
35) Dibromofluoromethane	8.165	113	122515	50.612	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.220%	
50) Toluene-d8	10.565	98	423897	49.287	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.580%	
62) 4-Bromofluorobenzene	12.847	95	130999	44.526	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	89.060%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
Data File : VN085521.D
Acq On : 27 Jan 2025 17:10
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 28 02:02:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

