

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
 Data File : VN085516.D
 Acq On : 27 Jan 2025 14:50
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 28 01:58:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.677	0.547	19.2	78	0.00
3 P	Chloromethane	0.733	0.639	12.8	85	0.00
4 C	Vinyl Chloride	0.737	0.633	14.1#	83	0.00
5 T	Bromomethane	0.445	0.432	2.9	93	0.00
6 T	Chloroethane	0.467	0.434	7.1	92	0.00
7 T	Trichlorofluoromethane	1.069	0.963	9.9	87	0.00
8 T	Diethyl Ether	0.369	0.399	-8.1	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.524	13.0	87	0.00
10 T	Methyl Iodide	0.690	0.742	-7.5	93	0.00
11 T	Tert butyl alcohol	0.092	0.124	-34.8#	121	0.00
12 CM	1,1-Dichloroethene	0.537	0.511	4.8#	87	0.00
13 T	Acrolein	0.126	0.178	-41.3#	115	0.00
14 T	Allyl chloride	0.871	0.887	-1.8	95	0.00
15 T	Acrylonitrile	0.294	0.339	-15.3	103	0.00
16 T	Acetone	0.261	0.281	-7.7	100	0.00
17 T	Carbon Disulfide	1.652	1.312	20.6	80	0.00
18 T	Methyl Acetate	0.793	0.949	-19.7	108	0.00
19 T	Methyl tert-butyl Ether	1.742	2.172	-24.7	105	0.00
20 T	Methylene Chloride	0.646	0.674	-4.3	97	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.556	3.0	90	0.00
22 T	Diisopropyl ether	1.933	2.261	-17.0	100	0.00
23 T	Vinyl Acetate	1.353	1.622	-19.9	98	0.00
24 P	1,1-Dichloroethane	1.178	1.235	-4.8	95	0.00
25 T	2-Butanone	0.384	0.433	-12.8	100	0.00
26 T	2,2-Dichloropropane	0.953	1.000	-4.9	96	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.728	-7.9	96	0.00
28 T	Bromochloromethane	0.548	0.503	8.2	84	0.00
29 T	Tetrahydrofuran	0.243	0.289	-18.9	100	0.00
30 C	Chloroform	1.218	1.284	-5.4#	99	0.00
31 T	Cyclohexane	1.024	0.824	19.5	85	0.00
32 T	1,1,1-Trichloroethane	1.068	1.050	1.7	93	0.00
33 S	1,2-Dichloroethane-d4	0.807	0.829	-2.7	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.347	0.327	5.8	88	0.00
36 T	1,1-Dichloropropene	0.487	0.434	10.9	90	0.00
37 T	Ethyl Acetate	0.491	0.517	-5.3	102	0.00
38 T	Carbon Tetrachloride	0.557	0.493	11.5	89	0.00
39 T	Methylcyclohexane	0.457	0.400	12.5	83	0.00
40 TM	Benzene	1.463	1.430	2.3	95	0.00
41 T	Methacrylonitrile	0.256	0.290	-13.3	103	0.00
42 TM	1,2-Dichloroethane	0.551	0.563	-2.2	99	0.00
43 T	Isopropyl Acetate	0.787	0.865	-9.9	102	0.00
44 TM	Trichloroethene	0.341	0.323	5.3	96	0.00
45 C	1,2-Dichloropropane	0.374	0.378	-1.1#	98	0.00
46 T	Dibromomethane	0.270	0.272	-0.7	98	0.00
47 T	Bromodichloromethane	0.549	0.583	-6.2	100	0.00
48 T	Methyl methacrylate	0.354	0.407	-15.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
 Data File : VN085516.D
 Acq On : 27 Jan 2025 14:50
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 28 01:58:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	95	0.00
50 S	Toluene-d8	1.232	1.150	6.7	87	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.523	-14.4	102	0.00
52 CM	Toluene	0.848	0.877	-3.4#	97	0.00
53 T	t-1,3-Dichloropropene	0.519	0.584	-12.5	101	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.613	-10.6	100	0.00
55 T	1,1,2-Trichloroethane	0.335	0.357	-6.6	101	0.00
56 T	Ethyl methacrylate	0.469	0.581	-23.9	102	0.00
57 T	1,3-Dichloropropane	0.583	0.632	-8.4	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.254	-19.2	101	0.00
59 T	2-Hexanone	0.321	0.375	-16.8	101	0.00
60 T	Dibromochloromethane	0.405	0.436	-7.7	101	0.00
61 T	1,2-Dibromoethane	0.334	0.357	-6.9	102	0.00
62 S	4-Bromofluorobenzene	0.422	0.421	0.2	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.341	0.320	6.2	95	0.00
65 PM	Chlorobenzene	1.095	1.094	0.1	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.414	-3.0	101	0.00
67 C	Ethyl Benzene	1.784	1.852	-3.8#	95	0.00
68 T	m/p-Xylenes	0.659	0.705	-7.0	95	0.00
69 T	o-Xylene	0.630	0.689	-9.4	97	0.00
70 T	Styrene	1.043	1.211	-16.1	99	0.00
71 P	Bromoform	0.288	0.332	-15.3	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.375	3.414	-1.2	95	0.00
74 T	N-amyl acetate	1.517	1.717	-13.2	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.192	1.227	-2.9	102	0.00
76 T	1,2,3-Trichloropropane	1.016	1.157	-13.9	124	0.00
77 T	Bromobenzene	0.882	0.908	-2.9	101	0.00
78 T	n-propylbenzene	3.995	4.099	-2.6	94	0.00
79 T	2-Chlorotoluene	2.586	2.619	-1.3	97	0.00
80 T	1,3,5-Trimethylbenzene	2.787	2.960	-6.2	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.423	-12.5	104	0.00
82 T	4-Chlorotoluene	2.574	2.680	-4.1	99	0.00
83 T	tert-Butylbenzene	2.341	2.385	-1.9	93	0.00
84 T	1,2,4-Trimethylbenzene	2.778	3.030	-9.1	96	0.00
85 T	sec-Butylbenzene	3.244	3.337	-2.9	93	0.00
86 T	p-Isopropyltoluene	2.631	2.795	-6.2	93	0.00
87 T	1,3-Dichlorobenzene	1.624	1.629	-0.3	99	0.00
88 T	1,4-Dichlorobenzene	1.683	1.636	2.8	100	0.00
89 T	n-Butylbenzene	2.327	2.427	-4.3	93	0.00
90 T	Hexachloroethane	0.618	0.576	6.8	95	0.00
91 T	1,2-Dichlorobenzene	1.620	1.617	0.2	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.231	-6.0	99	0.00
93 T	1,2,4-Trichlorobenzene	0.753	0.807	-7.2	99	0.00
94 T	Hexachlorobutadiene	0.400	0.362	9.5	94	0.00
95 T	Naphthalene	2.246	2.550	-13.5	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
Data File : VN085516.D
Acq On : 27 Jan 2025 14:50
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 28 01:58:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.812	-6.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6