

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
 Data File : VN085524.D
 Acq On : 27 Jan 2025 18:22
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 28 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.677	0.582	14.0	91	0.00
3 P	Chloromethane	0.733	0.618	15.7	89	0.00
4 C	Vinyl Chloride	0.737	0.644	12.6#	92	0.00
5 T	Bromomethane	0.445	0.380	14.6	89	0.00
6 T	Chloroethane	0.467	0.404	13.5	93	0.00
7 T	Trichlorofluoromethane	1.069	1.012	5.3	100	0.00
8 T	Diethyl Ether	0.369	0.352	4.6	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.584	3.0	106	0.00
10 T	Methyl Iodide	0.690	0.686	0.6	94	0.00
11 T	Tert butyl alcohol	0.092	0.106	-15.2	112	0.00
12 CM	1,1-Dichloroethene	0.537	0.522	2.8#	96	0.00
13 T	Acrolein	0.126	0.144	-14.3	101	0.00
14 T	Allyl chloride	0.871	0.936	-7.5	109	0.00
15 T	Acrylonitrile	0.294	0.311	-5.8	102	0.00
16 T	Acetone	0.261	0.266	-1.9	103	0.00
17 T	Carbon Disulfide	1.652	1.299	21.4	86	0.00
18 T	Methyl Acetate	0.793	0.885	-11.6	110	0.00
19 T	Methyl tert-butyl Ether	1.742	1.907	-9.5	100	0.00
20 T	Methylene Chloride	0.646	0.603	6.7	94	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.532	7.2	94	0.00
22 T	Diisopropyl ether	1.933	2.018	-4.4	97	0.00
23 T	Vinyl Acetate	1.353	1.467	-8.4	96	0.00
24 P	1,1-Dichloroethane	1.178	1.155	2.0	97	0.00
25 T	2-Butanone	0.384	0.419	-9.1	106	0.00
26 T	2,2-Dichloropropane	0.953	0.995	-4.4	104	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.666	1.3	96	0.00
28 T	Bromochloromethane	0.548	0.515	6.0	93	0.00
29 T	Tetrahydrofuran	0.243	0.272	-11.9	102	0.00
30 C	Chloroform	1.218	1.206	1.0#	101	0.00
31 T	Cyclohexane	1.024	0.882	13.9	98	0.00
32 T	1,1,1-Trichloroethane	1.068	1.063	0.5	103	0.00
33 S	1,2-Dichloroethane-d4	0.807	0.764	5.3	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.347	0.347	0.0	93	0.00
36 T	1,1-Dichloropropene	0.487	0.464	4.7	96	0.00
37 T	Ethyl Acetate	0.491	0.513	-4.5	101	0.00
38 T	Carbon Tetrachloride	0.557	0.554	0.5	100	0.00
39 T	Methylcyclohexane	0.457	0.457	0.0	95	0.00
40 TM	Benzene	1.463	1.461	0.1	97	0.00
41 T	Methacrylonitrile	0.256	0.279	-9.0	99	0.00
42 TM	1,2-Dichloroethane	0.551	0.561	-1.8	98	0.00
43 T	Isopropyl Acetate	0.787	0.858	-9.0	101	0.00
44 TM	Trichloroethene	0.341	0.333	2.3	99	0.00
45 C	1,2-Dichloropropane	0.374	0.378	-1.1#	98	0.00
46 T	Dibromomethane	0.270	0.264	2.2	95	0.00
47 T	Bromodichloromethane	0.549	0.575	-4.7	99	0.00
48 T	Methyl methacrylate	0.354	0.412	-16.4	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
 Data File : VN085524.D
 Acq On : 27 Jan 2025 18:22
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 28 02:04:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	107	0.00
50 S	Toluene-d8	1.232	1.251	-1.5	95	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.535	-17.1	104	0.00
52 CM	Toluene	0.848	0.901	-6.3#	99	0.00
53 T	t-1,3-Dichloropropene	0.519	0.568	-9.4	99	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.594	-7.2	97	0.00
55 T	1,1,2-Trichloroethane	0.335	0.354	-5.7	100	0.00
56 T	Ethyl methacrylate	0.469	0.564	-20.3	100	0.00
57 T	1,3-Dichloropropane	0.583	0.616	-5.7	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.220	-3.3	88	0.00
59 T	2-Hexanone	0.321	0.390	-21.5	105	0.00
60 T	Dibromochloromethane	0.405	0.427	-5.4	99	0.00
61 T	1,2-Dibromoethane	0.334	0.346	-3.6	100	0.00
62 S	4-Bromofluorobenzene	0.422	0.460	-9.0	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.341	0.333	2.3	103	0.00
65 PM	Chlorobenzene	1.095	1.080	1.4	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.401	0.2	102	0.00
67 C	Ethyl Benzene	1.784	1.894	-6.2#	101	0.00
68 T	m/p-Xylenes	0.659	0.725	-10.0	102	0.00
69 T	o-Xylene	0.630	0.696	-10.5	102	0.00
70 T	Styrene	1.043	1.196	-14.7	102	0.00
71 P	Bromoform	0.288	0.315	-9.4	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.375	3.596	-6.5	105	0.00
74 T	N-amyl acetate	1.517	1.643	-8.3	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.192	1.177	1.3	103	0.00
76 T	1,2,3-Trichloropropane	1.016	0.956	5.9	108	0.00
77 T	Bromobenzene	0.882	0.874	0.9	102	0.00
78 T	n-propylbenzene	3.995	4.279	-7.1	104	0.00
79 T	2-Chlorotoluene	2.586	2.643	-2.2	103	0.00
80 T	1,3,5-Trimethylbenzene	2.787	3.039	-9.0	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.407	-8.2	105	0.00
82 T	4-Chlorotoluene	2.574	2.645	-2.8	103	0.00
83 T	tert-Butylbenzene	2.341	2.514	-7.4	104	0.00
84 T	1,2,4-Trimethylbenzene	2.778	3.072	-10.6	102	0.00
85 T	sec-Butylbenzene	3.244	3.542	-9.2	104	0.00
86 T	p-Isopropyltoluene	2.631	2.940	-11.7	103	0.00
87 T	1,3-Dichlorobenzene	1.624	1.635	-0.7	105	0.00
88 T	1,4-Dichlorobenzene	1.683	1.597	5.1	103	0.00
89 T	n-Butylbenzene	2.327	2.444	-5.0	99	0.00
90 T	Hexachloroethane	0.618	0.598	3.2	104	0.00
91 T	1,2-Dichlorobenzene	1.620	1.578	2.6	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.231	-6.0	105	0.00
93 T	1,2,4-Trichlorobenzene	0.753	0.746	0.9	96	0.00
94 T	Hexachlorobutadiene	0.400	0.371	7.3	102	0.00
95 T	Naphthalene	2.246	2.331	-3.8	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012725\
Data File : VN085524.D
Acq On : 27 Jan 2025 18:22
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 28 02:04:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.762	0.727	4.6	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6