

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012819\
 Data File : VN053536.D
 Acq On : 28 Jan 2019 11:00
 Operator : JC/SP
 Sample : PB116640TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116640TB

Quant Time: Jan 29 06:57:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	826719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1322874	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1203899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	513110	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	428786	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
35) Dibromofluoromethane	7.59	113	417062	55.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.48%	
50) Toluene-d8	10.09	98	1605028	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
62) 4-Bromofluorobenzene	12.40	95	529704	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
Target Compounds						
16) Acetone	3.82	43	22196	11.578	ug/l	99
20) Methylene Chloride	4.56	84	14864	1.698	ug/l	91
43) Isopropyl Acetate	8.17	43	37360	2.998	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012819\
Data File : VN053536.D
Acq On : 28 Jan 2019 11:00
Operator : JC/SP
Sample : PB116640TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116640TB

Quant Time: Jan 29 06:57:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

