

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012819\
 Data File : VN053545.D
 Acq On : 28 Jan 2019 15:01
 Operator : JC/SP
 Sample : PB116640ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116640ZHE#01

Manual Integrations
 APPROVED

MMDadoda
 1/29/2019 9:52:51 AM

Quant Time: Jan 29 07:32:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	926773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1451454	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1281254	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	499933	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	470232	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
35) Dibromofluoromethane	7.59	113	453582	54.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.52%	
50) Toluene-d8	10.09	98	1753810	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
62) 4-Bromofluorobenzene	12.40	95	555046	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
Target Compounds						
16) Acetone	3.82	43	51543	23.983	ug/l	97
20) Methylene Chloride	4.56	84	33832	3.448	ug/l	99
25) 2-Butanone	6.84	43	14838	4.480	ug/l	97
29) Tetrahydrofuran	7.22	42	15007m	6.960	ug/l	
43) Isopropyl Acetate	8.17	43	53512	3.913	ug/l #	79
95) Naphthalene	15.14	128	7195	4.487	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012819\
Data File : VN053545.D
Acq On : 28 Jan 2019 15:01
Operator : JC/SP
Sample : PB116640ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116640ZHE#01

Manual Integrations
APPROVED

MMDadoda
1/29/2019 9:52:51 AM

Quant Time: Jan 29 07:32:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

