

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012819\
 Data File : VN053549.D
 Acq On : 28 Jan 2019 16:49
 Operator : JC/SP
 Sample : PB116640ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116640ZHE#05

Quant Time: Jan 29 08:30:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	921962	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1478200	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1354334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	585851	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	476037	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
35) Dibromofluoromethane	7.59	113	465132	55.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.26%	
50) Toluene-d8	10.09	98	1786894	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
62) 4-Bromofluorobenzene	12.40	95	603592	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
Target Compounds						
16) Acetone	3.82	43	45826	21.434	ug/l	96
20) Methylene Chloride	4.56	84	33538	3.436	ug/l	95
25) 2-Butanone	6.84	43	13062	3.965	ug/l	99
29) Tetrahydrofuran	7.22	42	13820	6.443	ug/l	98
43) Isopropyl Acetate	8.31	43	752097	54.004	ug/l #	51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012819\
Data File : VN053549.D
Acq On : 28 Jan 2019 16:49
Operator : JC/SP
Sample : PB116640ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116640ZHE#05

Quant Time: Jan 29 08:30:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

