

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012819\
 Data File : VN053550.D
 Acq On : 28 Jan 2019 17:16
 Operator : JC/SP
 Sample : PB116640ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116640ZHE#06

Quant Time: Jan 29 08:32:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	903262	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1444307	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1309745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	569251	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	468083	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
35) Dibromofluoromethane	7.59	113	458727	55.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.30%	
50) Toluene-d8	10.09	98	1733373	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
62) 4-Bromofluorobenzene	12.40	95	586565	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
Target Compounds						
16) Acetone	3.82	43	47097	22.485	ug/l	99
20) Methylene Chloride	4.56	84	33583	3.512	ug/l	97
25) 2-Butanone	6.84	43	11380	3.526	ug/l	97
29) Tetrahydrofuran	7.22	42	13225	6.294	ug/l	99
43) Isopropyl Acetate	8.31	43	482260	35.441	ug/l #	51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012819\
Data File : VN053550.D
Acq On : 28 Jan 2019 17:16
Operator : JC/SP
Sample : PB116640ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116640ZHE#06

Quant Time: Jan 29 08:32:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

