

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012819\
 Data File : VN053552.D
 Acq On : 28 Jan 2019 18:09
 Operator : JC/SP
 Sample : PB116640ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116640ZHE#08

Quant Time: Jan 29 08:29:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	847040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1373265	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1247495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	544044	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	437125	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
35) Dibromofluoromethane	7.59	113	425989	54.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.70%	
50) Toluene-d8	10.09	98	1655710	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	12.40	95	561880	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
Target Compounds						
16) Acetone	3.82	43	21712	11.054	ug/l	96
20) Methylene Chloride	4.55	84	14942	1.666	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012819\
Data File : VN053552.D
Acq On : 28 Jan 2019 18:09
Operator : JC/SP
Sample : PB116640ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116640ZHE#08

Quant Time: Jan 29 08:29:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

