

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012819\
 Data File : VN053555.D
 Acq On : 28 Jan 2019 19:30
 Operator : JC/SP
 Sample : K1011-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-01-012519-B

Quant Time: Jan 29 08:34:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	864762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1397487	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1249658	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	526584	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	450520	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
35) Dibromofluoromethane	7.59	113	438976	55.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.08%	
50) Toluene-d8	10.09	98	1672699	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
62) 4-Bromofluorobenzene	12.40	95	544606	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	
Target Compounds						
16) Acetone	3.82	43	32675	16.294	ug/l	99
20) Methylene Chloride	4.55	84	501558	54.782	ug/l	99
25) 2-Butanone	6.85	43	6381	2.065	ug/l	96
52) Toluene	10.16	92	7557	0.315	ug/l	89
95) Naphthalene	15.13	128	27661	5.941	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012819\
Data File : VN053555.D
Acq On : 28 Jan 2019 19:30
Operator : JC/SP
Sample : K1011-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OR-01-012519-B

Quant Time: Jan 29 08:34:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

