

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012819\
 Data File : VN053536.D
 Acq On : 28 Jan 2019 11:00
 Operator : JC/SP
 Sample : PB116640TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116640TB

Quant Time: Jan 29 06:57:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	826719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1322874	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1203899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	513110	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	428786	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
35) Dibromofluoromethane	7.59	113	417062	55.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.48%	
50) Toluene-d8	10.09	98	1605028	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
62) 4-Bromofluorobenzene	12.40	95	529704	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
Target Compounds						
16) Acetone	3.82	43	22196	11.578	ug/l	99
20) Methylene Chloride	4.56	84	14864	1.698	ug/l	91
43) Isopropyl Acetate	8.17	43	37360	2.998	ug/l #	84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012819\
Data File : VN053536.D
Acq On : 28 Jan 2019 11:00
Operator : JC/SP
Sample : PB116640TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116640TB

Quant Time: Jan 29 06:57:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN053536.D

Acq: 28 Jan 2019 11:00

Instrument :

MSVOA_N

ClientSampleId :


PB116640TB

Tot Ion:168 Resp: 826719

Ion Ratio Lower Upper

168 100

99 48.1 39.1 58.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67



300000

200000

100000

0

Time-->



#16

Acetone

Concen: 11.578 ug/l

RT: 3.82 min Scan# 636

Delta R.T. 0.01 min

Lab File: VN053536.D


Acq: 28 Jan 2019 11:00

Tgt Ion: 43 Resp: 22196

Ion Ratio Lower Upper

43 100

58 33.2 27.1 40.7



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO

3.82

8000


6000


4000

2000

0

Time-->









#20
Methvlene Chloride
Concen: 1.698 ug/l
RT: 4.56 min Scan# 866
Delta R.T. 0.01 min
Lab File: VN053536.D
Acq: 28 Jan 2019 11:00

Instrument :
MSVOA_N
ClientSampleId :
PB116640TB

Tot Ion: 84 Resp: 14864
Ion Ratio Lower Upper
84 100
49 101.9 94.3 141.5
51 33.2 29.4 44.2
86 64.7 51.4 77.0




Abundance Ion 83.90 (83.60 to 84.60): VNO
Ion 48.90 (48.60 to 49.60): VNO
Ion 50.90 (50.60 to 51.60): VNO
Ion 85.90 (85.60 to 86.60): VNO






#33
1,2-Dichloroethane-d4
Concen: Below ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN053536.D
Acq: 28 Jan 2019 11:00

Tgt Ion: 65 Resp: 428786
Ion Ratio Lower Upper
65 100
67 55.3 0.0 111.2



Abundance Ion 64.90 (64.60 to 65.60): VNO
Ion 66.90 (66.60 to 67.60): VNO







#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN053536.D
 Acq: 28 Jan 2019 11:00

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116640TB



Tot Ion:114 Resp: 1322874
 Ion Ratio Lower Upper
 114 100
 63 19.4 0.0 38.6
 88 15.1 0.0 30.4



Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V




Time--> 8.40 8.60 8.80




#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN053536.D
 Acq: 28 Jan 2019 11:00


Tgt Ion:113 Resp: 417062
 Ion Ratio Lower Upper
 113 100
 111 101.9 81.1 121.7
 192 23.1 18.2 27.2




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V



Time--> 7.50 7.60 7.70





#43

Isopropyl Acetate

Concen: 2.998 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN053536.D

Acq: 28 Jan 2019 11:00

Instrument :

MSVOA_N

ClientSampled :

PB116640TB


Tot Ion: 43 Resp: 37360

Ion Ratio Lower Upper

43 100

61 18.2 23.9 35.9#




87 11.7 11.0 16.6



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN053536.D


Acq: 28 Jan 2019 11:00

Tgt Ion: 98 Resp: 1605028

Ion Ratio Lower Upper



98 100


100 64.1 51.4 77.0



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN053536.D

Acq: 28 Jan 2019 11:00

Instrument :

MSVOA_N

ClientSampleId :

PB116640TB


Tot Ion: 95 Resp: 529704

Ion Ratio Lower Upper

95 100

174 94.9 0.0 183.2




176 90.4 0.0 177.4



Abundance Ion 94.90 (94.60 to 95.60): VN053399.D (-3290) (-)

Ion 173.80 (173.50 to 174.50): VN053399.D (-3290) (-)

Ion 175.80 (175.50 to 176.50): VN053399.D (-3290) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. 0.00 min

Lab File: VN053536.D

Acq: 28 Jan 2019 11:00


Tgt Ion: 117 Resp: 1203899

Ion Ratio Lower Upper

117 100

82 52.5 42.9 64.3



119 31.8 26.0 39.0




Abundance Ion 116.90 (116.60 to 117.60): VN053399.D (-2980) (-)

Ion 82.00 (81.70 to 82.70): VN053399.D (-2980) (-)

Ion 118.90 (118.60 to 119.60): VN053399.D (-2980) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN053536.D

Acq: 28 Jan 2019 11:00

Instrument :

MSVOA_N

ClientSampleId :

PB116640TB


Tot Ion:152 Resp: 513110

Ion Ratio Lower Upper

152 100

115 56.0 27.8 83.4

150 156.2 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

