

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012819\
 Data File : VN053552.D
 Acq On : 28 Jan 2019 18:09
 Operator : JC/SP
 Sample : PB116640ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116640ZHE#08

Quant Time: Jan 29 08:29:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	847040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1373265	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1247495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	544044	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	437125	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
35) Dibromofluoromethane	7.59	113	425989	54.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.70%	
50) Toluene-d8	10.09	98	1655710	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	12.40	95	561880	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
Target Compounds						
16) Acetone	3.82	43	21712	11.054	ug/l	96
20) Methylene Chloride	4.55	84	14942	1.666	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012819\
Data File : VN053552.D
Acq On : 28 Jan 2019 18:09
Operator : JC/SP
Sample : PB116640ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116640ZHE#08

Quant Time: Jan 29 08:29:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN053552.D

Acq: 28 Jan 2019 18:09

Instrument :

MSVOA_N

ClientSampleId :


PB116640ZHE#08

Tot Ion:168 Resp: 847040

Ion Ratio Lower Upper




168 100

99 49.2 39.1 58.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#16

Acetone

Concen: 11.054 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.01 min

Lab File: VN053552.D


Acq: 28 Jan 2019 18:09

Tgt Ion: 43 Resp: 21712

Ion Ratio Lower Upper



43 100


58 31.9 27.1 40.7



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO





#20

Methylene Chloride

Concen: 1.666 ug/l

RT: 4.55 min Scan# 864

Delta R.T. 0.00 min

Lab File: VN053552.D

Acq: 28 Jan 2019 18:09

Instrument :

MSVOA_N

ClientSampleId :

PB116640ZHE#08

Tot Ion: 84 Resp: 14942


Ion Ratio Lower Upper

84 100

49 120.7 94.3 141.5

51 42.1 29.4 44.2

86 67.0 51.4 77.0






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#33

1,2-Dichloroethane-d4

Concen: N.D. ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN053552.D


Acq: 28 Jan 2019 18:09

Tgt Ion: 65 Resp: 437125

Ion Ratio Lower Upper



65 100


67 55.5 0.0 111.2



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN053552.D

Acq: 28 Jan 2019 18:09

Instrument :

MSVOA_N

ClientSampleId :

PB116640ZHE#08


Tot Ion:114 Resp: 1373265

Ion Ratio Lower Upper

114 100

63 19.1 0.0 38.6




88 15.1 0.0 30.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN053552.D

Acq: 28 Jan 2019 18:09


Tgt Ion:113 Resp: 425989

Ion Ratio Lower Upper

113 100

111 102.9 81.1 121.7



192 23.5 18.2 27.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN053552.D

Acq: 28 Jan 2019 18:09

Instrument :

MSVOA_N

ClientSampleId :


PB116640ZHE#08

Tot Ion: 98 Resp: 1655710




Ion Ratio Lower Upper

98 100

100 64.9 51.4 77.0



Abundance Ion 98.00 (97.70 to 98.70): VN053399.D (-2569) (-)



#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN053552.D

Acq: 28 Jan 2019 18:09


Tgt Ion: 95 Resp: 561880

Ion Ratio Lower Upper



95 100


174 91.6 0.0 183.2

176 88.7 0.0 177.4



Abundance Ion 94.90 (94.60 to 95.60): VN053399.D (-3290) (-)





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. 0.00 min

Lab File: VN053552.D

Acq: 28 Jan 2019 18:09

Instrument :

MSVOA_N

ClientSampleId :

PB116640ZHE#08


Tot Ion:117 Resp: 1247495

Ion Ratio Lower Upper

117 100

82 52.6 42.9 64.3




119 32.2 26.0 39.0



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN053552.D

Acq: 28 Jan 2019 18:09


Tgt Ion:152 Resp: 544044

Ion Ratio Lower Upper

152 100

115 55.6 27.8 83.4

150 156.8 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

